

Configure screen savers

You must prevent your screen saver from launching during calibration and interfering with measurement.

1. Choose **Apple > System Settings**.
2. Go to **Apple > System Settings > Screen Saver**.
3. Select a screen saver.
4. Go to **Apple > System Settings > Lock Screen**.
5. Change the drop-down in **Start Screen Saver when inactive** to 30 minutes or longer.